



WASHINGTON STATE PATROL  
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: July 26, 2012 INITIALS: Dec OWNER: WASHINGTON STATE PATROL  
DISTRICT: 2 ADDRESS: I-5S MP141  
LOCATION: SeaTac SB 27 W2 COUNTY: King  
MAKE: Toledo TYPE: Trk Scale ACCURACY CLASS: III L  
MODEL: 8530 ID. NO: 4338198-4NY CAPACITY: 100000 lbs. SECTIONS: 2  
MINIMUM GRADUATIONS: 20 lbs. REG. ELEMENT: Digital PLATFORM 12x20

West2


LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR	LOAD POSITION	TEST WEIGHT	SCALE INDICATION	ERROR
SEC 1	20000	20000	0				
CENTER	20000	20000	0				
SEC 2	20000	20020	+20				

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as shown on the attached "Report of Calibration", Certification number(s): 200446-0-L3184-2 and 200446-0-L3184-4

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cromer Date: Aug 20, 2012

Print name: David Cromer



State of Washington  
County of Thurston  
Signed and sworn to (or affirmed before  
me on Aug. 20, 2012

Patricia A. Geppert  
(Signature of notary public)  
Patricia A. Geppert  
Printed Name of Notary  
My appointment expires Sept. 4, 2015